

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Adviser:

370

Virus Detection (Genome/Antigen) - Influenza A and B Viruses incl. Avian Influenza A Viruses

Influenza A viruses - qualitative

Name of test	Unit	Method	Manufacturer			
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
370126	positive (3)	positive (3)	SV		+	
370127	positive (3)	positive (3)	SV		+	
370128	below detection limit/negative (1)	below detection limit/negative (1)	SV		+	
370129	positive (3)	positive (3)	SV		+	
370130	below detection limit/negative (1)	below detection limit/negative (1)	SV		+	
370131	below detection limit/negative (1)	below detection limit/negative (1)	SV		+	

Subtyping of seasonal influenza A viruses

Name of test	Unit	Method	Manufacturer			
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
370126	negative for seasonal influenza A virus (1)	negative for seasonal influenza A virus (1)	ET		+	
370127	H3N2 (H3N2)	H3 (H3), H3N2 (H3N2)	ET		+	
370128	negative for seasonal influenza A virus (1)	negative for seasonal influenza A virus (1), not done (98)	ET		+	
370129	H1N1 pdm09 (H1N1)	H1 (H1), H1 pdm09 (H1), H1N1 (H1N1), H1N1 pdm09 (H1N1)	ET		+	
370130	negative for seasonal influenza A virus (1)	negative for seasonal influenza A virus (1), not done (98)	ET		+	
370131	negative for seasonal influenza A virus (1)	negative for seasonal influenza A virus (1), not done (98)	ET		+	

Qualitative detection of avian influenza A(H5Nx) virus with primers/probes specific for H5Nx or H5

Name of test	Unit	Method	Manufacturer			
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
370126	H5 (H5)	H5 (H5), H5N1 (H5N1)	ET		+	
370127	below detection limit/negative (1)	below detection limit/negative (1)	ET		+	
370128	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET		+	
370129	below detection limit/negative (1)	below detection limit/negative (1)	ET		+	
370130	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET		+	
370131	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET		+	

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Qualitative detection of influenza A(H1N1) pdm09 virus with primers/probes specific for H1N1 or H1

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	370126	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	370127	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	370128	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	370129	H1N1 pdm09 (H1N1)	H1 pdm09 (H1), H1N1 pdm09 (H1N1)	ET	+
	370130	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	370131	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	370126	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	370127	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	370128	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	370129	H1N1 pdm09 (H1N1)	H1 pdm09 (H1), H1N1 pdm09 (H1N1)	ET	+
	370130	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	370131	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+

Qualitative detection of avian influenza A(H7N9) virus with primers/probes specific for H7N9 or H7

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	370126	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	370127	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	370128	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	370129	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	370130	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	370131	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Influenza B viruses - qualitative

Name of test	Unit	Method	Manufacturer			
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	370126	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	370127	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	370128	positive (3)	positive (3)	SV		+
	370129	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	370130	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	370131	positive (3)	positive (3)	SV		+

371

Virus Genome Detection - Adenoviruses

Adenoviruses (DNA) - quantitative

Name of test	Unit	Method	Manufacturer				
	Copies/ml	PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	371085	2000000	149030	1490302600	ET	14903026	+
	371086	0.000	0.000	0.000	ET	0.000	+
	371087	82000	3039	30394000	ET	303940	+
	371088	250000	42284	422840500	ET	4228405	+

Adenoviruses (DNA) - qualitative

Name of test	Unit	Method	Manufacturer				
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria	
	371085	positive (3)	positive (3)	SV		+	
	371086	below detection limit/negative (1)	below detection limit/negative (1)	SV		+	
	371087	positive (3)	positive (3)	SV		+	
	371088	positive (3)	positive (3)	SV		+	
Name of test	Unit	Method	Manufacturer				
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria	
	371085	positive (3)	positive (3)	SV		+	
	371086	below detection limit/negative (1)	below detection limit/negative (1)	SV		+	
	371087	positive (3)	positive (3)	SV		+	
	371088	positive (3)	positive (3)	SV		+	

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Adenoviruses (DNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Viral gastroenteritis		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	371085	positive (3)	positive (3)	SV	+
	371086	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	371087	positive (3)	positive (3)	SV	+
	371088	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
FTD EPA		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	371085	positive (3)	positive (3)	SV	+
	371086	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	371087	positive (3)	positive (3)	SV	+
	371088	positive (3)	positive (3)	SV	+

Adenoviruses (DNA) - species identification

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	371085	species B (B)	species B (B)	ET	+
	371086	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	371087	species C (C)	species C (C)	ET	+
	371088	species D (D)	species D (D)	ET	+

Adenoviruses (DNA) - type identification

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	371085	AdV 11 (11)	AdV 11 (11)	ET	+
	371086	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	371087	AdV 2 (2)	AdV 2 (2)	ET	+
	371088	AdV 37 (37)	AdV 37 (37)	ET	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

340

Virus Genome Detection - Coronaviruses

MERS coronavirus (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	340075	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340077	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340079	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340080	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340081	below detection limit/negative (1)	below detection limit/negative (1)	SV	+

+

Sample not evaluated

Other human coronaviruses (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	340075	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340077	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340078	positive for HCoV 229E (229E)	positive for HCoV 229E (229E), positive for HCoV (for non differentiating HCoV tests) (Non)	SV	+
	340079	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340080	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340081	below detection limit/negative (1)			

+

Sample not evaluated

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

SARS-CoV-2 (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
NeuMoDx SARS-CoV-2 Assay		PCR / NAT (395)	NEUMODX (NE)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	340075	positive (3)	positive (3)	SV	+
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340077	positive (3)	positive (3)	SV	+
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340079	positive (3)	positive (3)	SV	+
	340080	positive (3)	positive (3)	SV	+
	340081	positive (3)	positive (3)	SV	+
Primer / Gene region	Sample	Your value			
	340075	N gene			
	340076	N gene			
	340077	N gene			
	340078	N gene			
	340079	N gene			
	340080	N gene			
	340081	N gene			
Name of test	Unit	Method	Manufacturer		
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	340075	positive (3)	positive (3)	SV	+
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340077	positive (3)	positive (3)	SV	+
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	340079	positive (3)	positive (3)	SV	+
	340080	positive (3)	positive (3)	SV	+
	340081	positive (3)	positive (3)	SV	+
Primer / Gene region	Sample	Your value			
	340075	RdRP gene			
	340076	RdRP gene			
	340077	RdRP gene			
	340078	RdRP gene			
	340079	RdRP gene			
	340080	RdRP gene			
	340081	RdRP gene			

Sample not evaluated

Sample not evaluated

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Name of test	Unit	Method	Manufacturer			
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	340075	positive (3)	positive (3)	SV		+
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	340077	positive (3)	positive (3)	SV		+
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	340079	positive (3)	positive (3)	SV		+
	340080	positive (3)	positive (3)	SV		+
	340081	positive (3)	positive (3)	SV		+
Primer / Gene region	Sample	Your value				
	340075	E gene				
	340076	E gene				
	340077	E gene				
	340078	E gene				
	340079	E gene				
	340080	E gene				
	340081	E gene				
Name of test	Unit	Method	Manufacturer			
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	340075	positive (3)	positive (3)	SV		+
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	340077	positive (3)	positive (3)	SV		+
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	340079	positive (3)	positive (3)	SV		+
	340080	positive (3)	positive (3)	SV		+
	340081	positive (3)	positive (3)	SV		+
Primer / Gene region	Sample	Your value				
	340075	N gene				
	340076	N gene				
	340077	N gene				
	340078	N gene				
	340079	N gene				
	340080	N gene				
	340081	N gene				

Sample not evaluated

Sample not evaluated

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Name of test	Unit	Method	Manufacturer			
RealStar SARS-CoV-2 RT-PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria	
	340075	positive (3)	positive (3)	SV	+	
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV	+	
	340077	positive (3)	positive (3)	SV	+	
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV	+	
	340079	positive (3)	positive (3)	SV	+	
	340080	positive (3)	positive (3)	SV	+	
	340081	positive (3)	positive (3)	SV	+	
Primer / Gene region	Sample	Your value				
	340075	E gene				
	340076	E gene				
	340077	E gene				
	340078	E gene				
	340079	E gene				
	340080	E gene				
	340081	E gene				
Name of test	Unit	Method	Manufacturer			
RealStar SARS-CoV-2 RT-PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria	
	340075	positive (3)	positive (3)	SV	+	
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV	+	
	340077	positive (3)	positive (3)	SV	+	
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV	+	
	340079	positive (3)	positive (3)	SV	+	
	340080	positive (3)	positive (3)	SV	+	
	340081	below detection limit/negative (1)	positive (3)	SV	+	
Primer / Gene region	Sample	Your value				
	340075	S gene				
	340076	S gene				
	340077	S gene				
	340078	S gene				
	340079	S gene				
	340080	S gene				
	340081	S gene				

Sample not evaluated

Sample not evaluated

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Name of test	Unit	Method	Manufacturer			
FTD SARS-CoV-2 assay		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	340075	positive (3)	positive (3)	SV		+
	340076	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	340077	positive (3)	positive (3)	SV		+
	340078	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	340079	positive (3)	positive (3)	SV		+
	340080	positive (3)	positive (3)	SV		+
	340081	positive (3)	positive (3)	SV		+
Primer / Gene region	Sample	Your value				
	340075	other				
	340076	other				
	340077	other				
	340078	other				
	340079	other				
	340080	other				
	340081	other				

Sample not evaluated

365

Virus Genome Detection - Cytomegalovirus Program 1

CMV (DNA) - quantitative						
Name of test	Unit	Method	Manufacturer			
	Copies/ml	PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value
	365169	0.000	0.000	0.000	SV	0.000
	365170	30000	11276	448901	SV	71146
	365171	170000	53041	2111587	SV	334664
	365172	4700	2053	81715	SV	12951

CMV (DNA) - qualitative						
Name of test	Unit	Method	Manufacturer			
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	365169	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	365170	positive (3)	positive (3)	SV		+
	365171	positive (3)	positive (3)	SV		+
	365172	positive (3)	positive (3)	SV		+



Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

372

Virus Genome Detection - Enteroviruses

Enteroviruses (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	372086	positive (3)	positive (3)	SV	+
	372087	positive (3)	positive (3)	SV	+
	372088	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	372089	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
RealStar Enterovirus RT-PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	372086	positive (3)	positive (3)	SV	+
	372087	positive (3)	positive (3)	SV	+
	372088	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	372089	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
FTD EPA		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	372086	positive (3)	positive (3)	SV	+
	372087	positive (3)	positive (3)	SV	+
	372088	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	372089	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	372086	positive (3)	positive (3)	SV	+
	372087	positive (3)	positive (3)	SV	+
	372088	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	372089	positive (3)	positive (3)	SV	+

Enteroviruses (RNA) - type identification

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)		
	372086	Coxsackievirus B3 (B3)	Coxsackievirus B3 (B3)	ET	+
	372087	Enterovirus 68 (68)	Enterovirus 68 (68)	ET	+
	372088	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	372089	Coxsackievirus B4 (B4)	Coxsackievirus B4 (B4)	ET	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann


Survey of 27 November 2020

376

Virus Genome Detection - Epstein Barr Virus

EBV (DNA) - quantitative

Name of test	Unit	Method	Manufacturer				
	Copies/ml	PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	376085	0.000	0.000	0.000	SV	0.000	+
	376086	3600	943	94300	SV	9430	+
	376087	8400	2072	207170	SV	20717	+
	376088	5000	1475	147470	SV	14747	+



EBV (DNA) - qualitative


Name of test	Unit	Method	Manufacturer				
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	376085	below detection limit/negative (1)	below detection limit/negative (1)		SV		+
	376086	positive (3)	positive (3)		SV		+
	376087	positive (3)	positive (3)		SV		+
	376088	positive (3)	positive (3)		SV		+

360

Virus Genome Detection - HIV-1 (RNA) Program 1

HIV-1 (RNA) - quantitative

Name of test	Unit	Method	Manufacturer				
Alinity m HIV-1 Assay	Copies/ml	PCR / NAT (395)	ABBOTT (AB)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	360169	0.000	0.000	0.000	SV	0.000	+
	360170	300	127	2006	SV	504	+
	360171	12000	4092	64856	SV	16291	+
	360172	33000	13112	207808	SV	52199	+
Name of test	Unit	Method	Manufacturer				
COBAS 4800 HIV-1 Test	Copies/ml	PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	360169	0.000	0.000	0.000	SV	0.000	+
	360170	400	127	2006	SV	504	+
	360171	23000	4092	64856	SV	16291	+
	360172	66000	13112	207808	SV	52199	+



HIV-1 (RNA) - qualitative

Name of test	Unit	Method	Manufacturer				
Alinity m HIV-1 Assay		PCR / NAT (395)	ABBOTT (AB)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	360169	below detection limit/negative (1)	below detection limit/negative (1)		SV		+
	360170	positive (3)	positive (3)		SV		+
	360171	positive (3)	positive (3)		SV		+
	360172	positive (3)	positive (3)		SV		+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

HIV-1 (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
COBAS 4800 HIV-1 Test		PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	360169	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	360170	positive (3)	positive (3)	SV	+
	360171	positive (3)	positive (3)	SV	+
	360172	positive (3)	positive (3)	SV	+

377

Virus Genome Detection - Hepatitis A Virus

HAV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
RealStar HAV RT-PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	377169	positive (3)	positive (3)	SV	+
	377170	positive (3)	positive (3)	SV	+
	377171	positive (3)	positive (3)	SV	+
	377172	below detection limit/negative (1)	below detection limit/negative (1)	SV	+

396

Virus Genome Detection - Hepatitis B Virus Genotyping

HBV - genotyping

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)		
	396021	Genotype A (A)	Genotype A (A)	ET	+
	396022	Genotype B (B)	Genotype B (B)	ET	+
	396023	Genotype D (D)	Genotype D (D)	ET	+
	396024	Genotype A (A)	Genotype A (A)	ET	+

HBV - genosubtyping

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)		
	396021	Genosubtype A1 (A1)	Genosubtype A1 (A1)	ET	+
	396022	Genosubtype B2 (B2)	Genosubtype B2 (B2)	ET	+
	396023	Genosubtype D1 (D1)	Genosubtype D1 (D1)	ET	+
	396024	Genosubtype A2 (A2)	Genosubtype A2 (A2)	ET	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

361

Virus Genome Detection - Hepatitis B Virus Program 1

HBV (DNA) - quantitative

Name of test	Unit	Method	Manufacturer				
Alinity m HBV Assay	IU/ml	PCR / NAT (395)	ABBOTT (AB)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	361169	7600	2017	31972	SV	8031	+
	361170	46000	12097	191720	SV	48158	+
	361171	2100	771	12218	SV	3069	+
	361172	23000	5939	94128	SV	23644	+
Name of test	Unit	Method	Manufacturer				
COBAS 4800 HBV Test	IU/ml	PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	361169	7900	2017	31972	SV	8031	+
	361170	59000	12097	191720	SV	48158	+
	361171	3100	771	12218	SV	3069	+
	361172	28000	5939	94128	SV	23644	+

HBV (DNA) - qualitative

Name of test	Unit	Method	Manufacturer				
Alinity m HBV Assay		PCR / NAT (395)	ABBOTT (AB)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	361169	positive (3)	positive (3)		SV		+
	361170	positive (3)	positive (3)		SV		+
	361171	positive (3)	positive (3)		SV		+
	361172	positive (3)	positive (3)		SV		+
Name of test	Unit	Method	Manufacturer				
COBAS 4800 HBV Test		PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	361169	positive (3)	positive (3)		SV		+
	361170	positive (3)	positive (3)		SV		+
	361171	positive (3)	positive (3)		SV		+
	361172	positive (3)	positive (3)		SV		+

362

Virus Genome Detection - Hepatitis C Virus Program 1

HCV (RNA) - quantitative

Name of test	Unit	Method	Manufacturer				
COBAS 4800 HCV Test	IU/ml	PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	362169	0.000	0.000	0.000	SV	0.000	+
	362170	2500	1459	23126	SV	5809	+
	362171	320	155	2460	SV	618	+
	362172	1100	506	8014	SV	2013	+


Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

HCV (RNA) - quantitative

Name of test	Unit	Method	Manufacturer				
Alinity m HCV Assay	IU/ml	PCR / NAT (395)	ABBOTT (AB)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	362169	0.000	0.000	0.000	SV	0.000	+
	362170	4800	1459	23126	SV	5809	+
	362171	490	155	2460	SV	618	+
	362172	2500	506	8014	SV	2013	+



HCV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer				
Abbott RealTime HCV		PCR / NAT (395)	ABBOTT (AB)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type			Meets criteria
	362169	below detection limit/negative (1) positive (3)	below detection limit/negative (1) positive (3)	SV			+
	362170	positive (3)	positive (3)	SV			+
	362171	positive (3)	positive (3)	SV			+
	362172	positive (3)	positive (3)	SV			+

Name of test	Unit	Method	Manufacturer				
COBAS 4800 HCV Test		PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type			Meets criteria
	362169	below detection limit/negative (1)	below detection limit/negative (1)	SV			+
	362170	positive (3)	positive (3)	SV			+
	362171	positive (3)	positive (3)	SV			+
	362172	positive (3)	positive (3)	SV			+

Name of test	Unit	Method	Manufacturer				
Alinity m HCV Assay		PCR / NAT (395)	ABBOTT (AB)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type			Meets criteria
	362169	below detection limit/negative (1)	below detection limit/negative (1)	SV			+
	362170	positive (3)	positive (3)	SV			+
	362171	positive (3)	positive (3)	SV			+
	362172	positive (3)	positive (3)	SV			+



Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann


Survey of 27 November 2020

380

Virus Genome Detection - Hepatitis E Virus

HEV (RNA) - quantitative

Name of test	Unit	Method	Manufacturer				
RealStar HEV RT-PCR Kit 2.0	IU/ml	PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	380077	64000	6035	603470	ET	60347	+
	380078	1400	296	29580	ET	2958	+
	380079	220000	17560	1755960	ET	175596	+
	380080	51000	4119	411900	ET	41190	+
Name of test	Unit	Method	Manufacturer				
RealStar HEV RT-PCR Kit 2.0	IU/ml	PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	380077	49000	6035	603470	ET	60347	+
	380078	1000	296	29580	ET	2958	+
	380079	150000	17560	1755960	ET	175596	+
	380080	48000	4119	411900	ET	41190	+



HEV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer				
RealStar HEV RT-PCR Kit 2.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	380077	positive (3)	positive (3)		SV		+
	380078	positive (3)	positive (3)		SV		+
	380079	positive (3)	positive (3)		SV		+
	380080	positive (3)	positive (3)		SV		+
Name of test	Unit	Method	Manufacturer				
RealStar HEV RT-PCR Kit 2.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)				
Result (qual)	Sample	Your specification(s)	Correct specification(s)		TV-Type		Meets criteria
	380077	positive (3)	positive (3)		SV		+
	380078	positive (3)	positive (3)		SV		+
	380079	positive (3)	positive (3)		SV		+
	380080	positive (3)	positive (3)		SV		+




363

Virus Genome Detection - Herpes Simplex Virus Type 1/Type 2

HSV 1 (DNA) - quantitative

Name of test	Unit	Method	Manufacturer				
RealStar HSV PCR Kit 1.0	Copies/ml	PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	363127	0.000	0.000	0.000	SV	0.000	+
	363128	30000	3960	396040	SV	39604	+
	363129	110000	12853	1285270	SV	128527	+
	363130	0.000	0.000	0.000	SV	0.000	+
	363131	0.000	0.000	0.000	SV	0.000	+
	363132	10000	1376	137600	SV	13760	+



Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

HSV 1 (DNA) - qualitative

Name of test	Unit	Method	Manufacturer			
RealStar alpha Herpesvirus PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	363127	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	363128	positive (3)	positive (3)	SV		+
	363129	positive (3)	positive (3)	SV		+
	363130	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	363131	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	363132	positive (3)	positive (3)	SV		+

HSV 2 (DNA) - quantitative

Name of test	Unit	Method	Manufacturer			
RealStar HSV PCR Kit 1.0	Copies/ml	PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value
	363127	42000	3936	393570	SV	39357
	363128	0.000	0.000	0.000	SV	0.000
	363129	0.000	0.000	0.000	SV	0.000
	363130	0.000	0.000	0.000	SV	0.000
	363131	14000	1411	141080	SV	14108
	363132	0.000	0.000	0.000	SV	0.000

HSV 2 (DNA) - qualitative

Name of test	Unit	Method	Manufacturer			
RealStar alpha Herpesvirus PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	363127	positive (3)	positive (3)	SV		+
	363128	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	363129	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	363130	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	363131	positive (3)	positive (3)	SV		+
	363132	below detection limit/negative (1)	below detection limit/negative (1)	SV		+

HSV (DNA) - qualitative without differentiation of HSV 1 and HSV 2

Name of test	Unit	Method	Manufacturer			
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	363127	positive (3)	positive (3)	SV		+
	363128	positive (3)	positive (3)	SV		+
	363129	positive (3)	positive (3)	SV		+
	363130	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	363131	positive (3)	positive (3)	SV		+
	363132	positive (3)	positive (3)	SV		+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

385

Virus Genome Detection - Human Metapneumovirus

HMPV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	385053	positive (3)	positive (3)	SV	+
	385054	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	385055	positive (3)	positive (3)	SV	+
	385056	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	385053	positive (3)	positive (3)	SV	+
	385054	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	385055	positive (3)	positive (3)	SV	+
	385056	positive (3)	positive (3)	SV	+

373

Virus Genome Detection - Human Papilloma Viruses

Differentiation of "High Risk" and "Low Risk" types of HPV

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
INNO-LiPA HPV Genotyping Extra II		PCR / NAT (395)	FUJIREBIO (FJ)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	373106	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	373107	positive for High Risk-HPV (3)	positive for High Risk-HPV (3)	SV	+
	373108	positive for High Risk-HPV (3)	positive for High Risk-HPV (3)	SV	+
	373109	positive for High Risk-HPV (3)	positive for High Risk-HPV (3)	SV	+
	373110	positive for Low Risk-HPV (4)	positive for Low Risk-HPV (4)	SV	+

Detection of "High Risk" HPV (without detection of "Low Risk" HPV)

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
Cobas 4800 HPV Test		PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	373106	below detection limit/negative for High Risk-HPV (1)	below detection limit/negative for High Risk-HPV (1)	SV	+
	373107	positive for High Risk-HPV (3)	positive for High Risk-HPV (3)	SV	+
	373108	positive for High Risk-HPV (3)	positive for High Risk-HPV (3)	SV	+
	373109	positive for High Risk-HPV (3)	positive for High Risk-HPV (3)	SV	+
	373110	below detection limit/negative for High Risk-HPV (1)	below detection limit/negative for High Risk-HPV (1)	SV	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Typing of "High Risk" HPV and "Low Risk" HPV

Name of test	Unit	Method	Manufacturer			
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)			
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	373106	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET		+
	373107	HPV 16 (High Risk) (16)	HPV 16 (High Risk) (16)	ET		+
	373108	HPV 16 (High Risk) (16)	HPV 16 (High Risk) (16)	ET		+
	373109	HPV 18 (High Risk) (18)	HPV 18 (High Risk) (18)	ET		+
	373110	HPV 6 (Low Risk) (6)	HPV 6 (Low Risk) (6), HPV 6, HPV 11 (Low Risk) (6,11)	ET		+
Name of test	Unit	Method	Manufacturer			
INNO-LiPA HPV Genotyping Extra II		PCR / NAT (395)	FUJIREBIO (FJ)			
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	373106	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET		+
	373107	HPV 16 (High Risk) (16)	HPV 16 (High Risk) (16)	ET		+
	373108	HPV 16 (High Risk) (16)	HPV 16 (High Risk) (16)	ET		+
	373109	HPV 18 (High Risk) (18)	HPV 18 (High Risk) (18)	ET		+
	373110	HPV 6 (Low Risk) (6)	HPV 6 (Low Risk) (6), HPV 6, HPV 11 (Low Risk) (6,11)	ET		+

Typing of "High Risk" HPV (without detection of "Low Risk" HPV)

Name of test	Unit	Method	Manufacturer			
Cobas 4800 HPV Test		PCR / NAT (395)	ROCHE DIAGNOSTICS (RO)			
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	373106	negative for High Risk HPV (1)	negative for High Risk HPV (1), not done (98)	ET		+
	373107	HPV 16 (High Risk) (16)	HPV 16 (High Risk) (16)	ET		+
	373108	HPV 16 (High Risk) (16)	HPV 16 (High Risk) (16)	ET		+
	373109	HPV 18 (High Risk) (18)	HPV 18 (High Risk) (18), HPV 18, HPV 45 (High Risk) (18,45)	ET		+
	373110	negative for High Risk HPV (1)	negative for High Risk HPV (1), not done (98)	ET		+

393

Virus Genome Detection - Human Rhinoviruses

HRV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer			
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	393045	positive (3)	positive (3)	SV		+
	393046	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	393047	positive (3)	positive (3)	SV		+
	393048	positive (3)	positive (3)	SV		+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

HRV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Respiratory pathogens 21		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	393045	positive (3)	positive (3)	SV	+
	393046	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	393047	positive (3)	positive (3)	SV	+
	393048	positive (3)	positive (3)	SV	+

HRV (RNA) - type identification

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)		
	393045	HRV A30 (A30)	HRV A30 (A30)	ET	+
	393046	below detection limit/negative (1)	below detection limit/negative (1)	ET	+
	393047	HRV A1 (A1)	HRV A1 (A1)	ET	+
	393048	HRV A30 (A30)	HRV A30 (A30)	ET	+

386

Virus Genome Detection - Measles Virus

Measles virus (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	386053	positive (3)	positive (3)	SV	+
	386054	positive (3)	positive (3)	SV	+
	386055	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	386056	positive (3)	positive (3)	SV	+

Measles virus (RNA) - type identification

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)		
	386053	Genotype B3 (B3)	Genotype B3 (B3)	ET	+
	386054	Genotype H1 (H1)	Genotype H1 (H1)	ET	+
	386055	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	386056	Genotype D8 (D8)	Genotype D8 (D8)	ET	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

387

Virus Genome Detection - Mumps Virus

Mumps virus (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	387049	positive (3)	positive (3)	SV	+
	387050	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	387051	positive (3)	positive (3)	SV	+
	387052	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	387049	positive (3)	positive (3)	SV	+
	387050	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	387051	positive (3)	positive (3)	SV	+
	387052	positive (3)	positive (3)	SV	+

Mumps virus (RNA) - type identification

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	387049	Genotype G (G)	Genotype G (G)	ET	+
	387050	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	387051	Genotype G (G)	Genotype G (G)	ET	+
	387052	Genotype C (C)	Genotype C (C)	ET	+

381

Virus Genome Detection - Norovirus

Norovirus (RNA) genogroup I - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Viral gastroenteritis		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	381062	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	381063	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	381064	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	381065	below detection limit/negative (1)	below detection limit/negative (1)	SV	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

Norovirus (RNA) genogroup II - qualitative

Name of test	Unit	Method	Manufacturer			
FTD Viral gastroenteritis		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	381062	positive (3)	positive (3)	SV		+
	381063	positive (3)	positive (3)	SV		+
	381064	positive (3)	positive (3)	SV		+
	381065	below detection limit/negative (1)	below detection limit/negative (1)	SV		+

367

Virus Genome Detection - Parvovirus B19

Parvovirus B19 (DNA) - quantitative

Name of test	Unit	Method	Manufacturer				
RealStar Parvovirus B19 Kit 1.0	IU/ml	PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)				
Result (quant)	Sample	Your value	Lower limit	Upper limit	TV-Type	Target value	Meets criteria
	367169	360000	193123	7688348	ET	1218521	+
	367170	0.000	0.000	0.000	ET	0.000	+
	367171	1800	1696	67519	ET	10701	+
	367172	51000	41655	1658320	ET	262826	+

Parvovirus B19 (DNA) - qualitative

Name of test	Unit	Method	Manufacturer			
RealStar Parvovirus B19 Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	367169	positive (3)	positive (3)	SV		+
	367170	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	367171	positive (3)	positive (3)	SV		+
	367172	positive (3)	positive (3)	SV		+

359

Virus Genome Detection - Respiratory Syncytial Virus (Antigen/Genome)

RSV antigen

Name of test	Unit	Method	Manufacturer			
Light Diagnostics Respiratory Panel I Viral Screening and Identification IFA		Immuno fluorescence (240)	MILLIPORE (ML)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	359065	positive (3)	positive (3)	SV		+
	359066	positive (3)	positive (3)	SV		+
	359067	negative (1)	negative (1)	SV		+
	359068	positive (3)	positive (3)	SV		+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

RSV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Respiratory pathogens 21		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	359065	positive (3)	positive (3)	SV	+
	359066	positive (3)	positive (3)	SV	+
	359067	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	359068	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
RealStar RSV RT-PCR Kit 3.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	359065	positive (3)	positive (3)	SV	+
	359066	positive (3)	positive (3)	SV	+
	359067	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	359068	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	359065	positive (3)	positive (3)	SV	+
	359066	positive (3)	positive (3)	SV	+
	359067	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	359068	positive (3)	positive (3)	SV	+

401

Virus Genome Detection - Rotaviruses

Rotaviruses (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
FTD Viral gastroenteritis		PCR / NAT (395)	FAST-TRACK DIAGNOSTICS (FT)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	401045	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	401046	positive (3)	positive (3)	SV	+
	401047	positive (3)	positive (3)	SV	+
	401048	positive (3)	positive (3)	SV	+

Listing and evaluation of the results

1: Dr. med. Max Mustermann
Laboratory Mustermann

Survey of 27 November 2020

389

Virus Genome Detection - Rubella Virus

Rubella virus (RNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)		
	389049	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	389050	positive (3)	positive (3)	SV	+
	389051	positive (3)	positive (3)	SV	+
	389052	positive (3)	positive (3)	SV	+
Name of test	Unit	Method	Manufacturer		
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	389049	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	389050	positive (3)	positive (3)	SV	+
	389051	positive (3)	positive (3)	SV	+
	389052	positive (3)	positive (3)	SV	+

Rubella virus (RNA) - type identification

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
		PCR / NAT (395)	IN-HOUSE PRODUCTION (ZX)		
Typing result	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	389049	below detection limit/negative (1)	below detection limit/negative (1), not done (98)	ET	+
	389050	Genotype 1G (1G)	Genotype 1G (1G)	ET	+
	389051	Genotype 1E (1E)	Genotype 1E (1E)	ET	+
	389052	Genotype 2B (2B)	Genotype 2B (2B)	ET	+

366

Virus Genome Detection - Varicella Zoster Virus

VZV (DNA) - quantitative

Name of test	Unit	Method	Manufacturer	TV-Type	Target value	Meets criteria
RealStar VZV PCR Kit 1.0	Copies/ml	PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (quant)	Sample	Your value	Lower limit	Upper limit		
	366085	1500000	125424	1254230	SV	1254243
	366086	520000	39534	3953380	SV	395338
	366087	0.000	0.000	0.000	SV	0.000
	366088	56000	4935	493470	SV	49347

VZV (DNA) - qualitative

Name of test	Unit	Method	Manufacturer	TV-Type	Meets criteria
RealStar alpha Herpesvirus PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)		
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type	Meets criteria
	366085	positive (3)	positive (3)	SV	+
	366086	positive (3)	positive (3)	SV	+
	366087	below detection limit/negative (1)	below detection limit/negative (1)	SV	+
	366088	positive (3)	positive (3)	SV	+

Listing and evaluation of the results

 1: Dr. med. Max Mustermann
 Laboratory Mustermann

Survey of 27 November 2020

391

Virus Genome Detection - West Nile Virus
WNV (RNA) - qualitative

Name of test	Unit	Method	Manufacturer			
RealStar WNV RT-PCR Kit 1.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	391101	positive (3)	positive (3)	SV		+
	391102	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	391103	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	391104	positive (3)	positive (3)	SV		+
	391105	positive (3)	positive (3)	SV		+
	391106	positive (3)	positive (3)	SV		+
Name of test	Unit	Method	Manufacturer			
RealStar WNV RT-PCR Kit 2.0		PCR / NAT (395)	ALTONA DIAGNOSTIC (AK)			
Result (qual)	Sample	Your specification(s)	Correct specification(s)	TV-Type		Meets criteria
	391101	positive (3)	positive (3)	SV		+
	391102	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	391103	below detection limit/negative (1)	below detection limit/negative (1)	SV		+
	391104	positive (3)	positive (3)	SV		+
	391105	positive (3)	positive (3)	SV		+
	391106	positive (3)	positive (3)	SV		+